

NATIONAL INSTITUTE OF UNANI MEDICINE

Kottigepalya, Magadi Main Road, Bangalore – 560 091. (An autonomous body under Govt. of India, Ministry of AYUSH) Tel : 080-23584

F.No.04-03/2017-18/NIUM/Admn./Vol.IV/2956 (Procurement of Hospital Equipment) **To,** Tel : 080-23584260 Fax: 080 -23584180 Date. 11/10/2018.

Subject: Invitation of re-quotation for supply of Striker Motorized Bed for Recovery Room, in Hospital, NIUM: – reg. *********

Sir,

With reference to the subject mentioned above, Quotations are invited from reputed registered dealers for the supply of instruments as detailed below.

Sl. No	Name of the Items	Specification /Make	Quantity Required.	Amount	Total GST	Grant Total
INO	Items		Kequirea.		Amount	Total
				()	(`)	(`)
01	Striker	Model No SV1 Versatile Patient care Bed	02 Sets.		()	()
01	Motorized Bed	Battery back up, DC motor for quiet operation	02 Sets.			
	including	Manual CPR on both side for Backrest				
	mattress	Lockable and Tuck Away siderails				
	mattress	Removable Litter cover for cleaning				
		Removable Headboard and Footboard				
		Central and Directional Lockable casters				
		One antistatic Caster to ground Frame				
		Plastic Crash Bumpers				
		Easy adjustable foot prop for Knee-bend				
		Positioning and foot elevation				
		Electro Static Painted metal Frame				
		Angle indicator for Backrest				
		Angle indicator for bed angle				
		IP X 4 Liquid splash protection				
		Dual wheel caster from tente				
		2 stage stainless steel IV pole				
		Upright oxygen bottle holder				
		Outer controls for caregiver convenience				
		Inner controlfor patient comfort				
		Lockable and Tuck Away functionally				
		Xray Transluscent				
		Nurse control pendent				
		Automatic contour position				
		One botton position for Cardiac, egress and				
		vascular				
		Examnination, Trendelburg & Reverse				
		Trendelburg				

LED Light indicator for motion lockout, power concentration and batter charge. Head of Beg 30 deg Dual wheel – caster from Tente Colour – Blue		
Mattress: four way stretch cover Cover material is POLYURETHANE coated polyster Fire retardant mattress system Surface performance well on testing that stimulated a lifetime of boosting, a lifetime of		
patient ingress/ egress, and a lifetime of bed flexing.		

The quotations may be submitted in a sealed cover quoting the competitive rate with above specification by clearly mentioning taxes applicable if any., time required for supplying the items with the terms & conditions, Copy of GST

Wherever required Test Certificate should be sent along with relevant document.

Incomplete bids are liable for rejection. Late / delayed offer will not be considered.

NIUM reserves the right to reject any or all tender without assigning any reason.

The envelope containing quotation should be super scribed as "Re-Quotation for supply

of Striker Motorized Bed for Recovery Room in Hospital, NIUM"

The last date for the submission of quotation is: <u>30/10/2018.</u>

Yours faithfully,

Sd/-

(Suresha) Accounts Officer

Distribution : 1.Dr.S.Shah Alam, DMS, for information. 2.Website in-charge for information and to upload quotation in NIUM- website. 3.Concern file. 4.Guard file.